

CONFIRMATION

Notification: 0000001013_07_03
on changes according to EN 15267 regarding certificate 0000001013_07 dated 29 July 2022

AMS designation: CEMS II e for CO, NO, NO₂, N₂O, SO₂, HCl, HF, NH₃, CH₄, CH₂O, O₂, H₂O and CO₂

Manufacturer: Gasmot Technologies Oy
Mestarintie 6
01730 Vantaa
Finland

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 05 July 2023
Publication: BAnz AT 02.08.2023 B7**

III. Notifications as regards the uniform practice in monitoring emissions and ambient air:

7 Notification as regards Federal Environment Agency (UBA) notices of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter I number 3.3) and of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 26)

The oxygen measuring module OXITEC 500E SME5 from the CEMS II e measuring system for the components O₂, CO, NO, NO₂, N₂O, SO₂, HCl, HF, NH₃, H₂O, CO₂, H₂CO and CH₄ from Gasmot Technology Oy can now also be equipped with the display board Rev. 10e. For additional implementation of the display board Rev. 10e the software of the oxygen measuring module has been updated.


The current software versions of the CEMS II e measuring system are now:


Calcmet: 12.240 with evaluation module 4.42.2

OXITEC 500E: 4.14.

Statement issued by TÜV Rheinland Energy GmbH dated 31 March 2023

TÜV Rheinland Energy GmbH
Cologne, 02 August 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.