

CONFIRMATION

Notification: 0000025926_05_01
on changes according to EN 15267 regarding certificate 0000025926_05 dated 12 February 2020

AMS designation: MCS 100 FT for CO, NO, NO₂, N₂O, SO₂, HCl, HF, NH₃, CH₄, TOC, O₂, H₂O and CO₂

Manufacturer: SICK AG
Nimburger Straße 11
79276 Reute
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 31 March 2021
Publication: BAnz AT 03.05.2021 B9**


**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air**


**53 Notification as regards Federal Environment Agency (UBA) notices
of 10 January 2011 (BAnz. p. 294, chapter I, number 3.1) and
of 3 July 2018 (BAnz AT 17.07.2018 B9, chapter III notification 25)**

In the case of the MCS100FT measuring system for O₂, CO, SO₂, NO, NO₂, HCl, HF, CH₄, CO₂, H₂O, N₂O, NH₃ and TOC manufactured by SICK AG, the soldering point for electrically conducting the suction voltage at the FID has so far been made of an Fe/Ni alloy. In the future, this assembly can also be made of stainless steel.

Statement issued by TÜV Rheinland Energy GmbH dated 18 September 2020

TÜV Rheinland Energy GmbH
Cologne, 25 May 2021


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.