

CONFIRMATION

Notification: 0000025930_04_02
on changes according to EN 15267 regarding certificate 0000025930_04 dated 12 February 2020

AMS designation: PowerCEMS100 for CO, NO, NO₂, N₂O, SO₂, CH₄, O₂ and CO₂

Manufacturer: SICK AG
Nimburger Straße 11
79276 Reute
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 31 March 2021
Publication: BAnz AT 03.05.2021 B9**


**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air**


**56 Notification as regards Federal Environment Agency (UBA) notices
of 23 February 2012 (BAnz. p. 920, chapter I number 5.1) and
of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter IV notification 59)**

In future the PowerCEMS100 modular measuring system manufactured by SICK AG can be equipped with a housing flushing by means of inert gas or clean air for the measuring module GMS800 DEFOR when installed in areas with polluted ambient air. This flushing can be done either for the whole 19" housing or for the filter wheel housing. Corresponding constructive adjustments were made to the housings for the connection of the coil air.

Statement issued by TÜV Rheinland Energy GmbH dated 18 September 2020

TÜV Rheinland Energy GmbH
Cologne, 25 May 2021


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.