

CONFIRMATION

Notification: 0000032301_02_01
on changes according to EN 15267 regarding certificate 0000032301_02 dated 02 March 2023

AMS designation: PG 350 E for CO, NO_x, SO₂, O₂ and CO₂

Manufacturer: HORIBA Europe GmbH
Hans-Mess-Str. 6
61440 Oberursel /Ts.
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**30 Notification as regards Federal Environment Agency (UBA) notices
of 12 February 2013 (BAnz AT 05.03.2013 B10, chapter I number 5.2) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 40)**


The current software version of the PG 350E measuring system for NO_x, SO₂, CO, CO₂
and O₂ from the company HORIBA Europe GmbH is:


P2000788001H/1.27

The measuring system is also suitable for use at plants with regard to the
44th BImSchV.

Statement issued by TÜV Rheinland Energy GmbH dated 17 September 2022 (software
version) and dated 10 January 2023 (44th BImSchV)

TÜV Rheinland Energy GmbH
Cologne, 25 April 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.