

# CONFIRMATION

Notification: 0000034862\_03\_01  
on changes according to EN 15267 regarding certificate 0000034862\_03 dated 05 November 2019

---

**AMS designation:** GC 5000 BTX version FID for benzene

**Manufacturer:** AMA Instruments GmbH  
Lise-Meitner-Strasse 8  
89081 Ulm  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 28 June 2022  
Publication: BAnz AT 28.07.2022 B4**

**III. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**37 Notification as regards Federal Environment Agency (UBA) notice  
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter III number 3.1)**

The current software versions of the measuring device gas chromatograph GC 5000  
BTX version FID for benzene of AMA Instruments GmbH are:

Common.Control v. 1.1  
SS.Control v. 2.1  
AMA\_PEAK.log v. 1.3


Various hardware changes have been made to the measuring equipment. These  
concern the following points:


- new connector modules at the back of the device
- new electronic pressure controller
- optimization of valve drive 6-port valve
- new actuator for furnace damper
- minor housing adaptations

Statement issued by TÜV Rheinland Energy GmbH dated 26 April 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 01 August 2022

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.