

CONFIRMATION

Notification: 0000035010_02_02
on changes according to EN 15267 regarding certificate 0000035010_02 dated 16 February 2022

AMS designation: ENDA-5000 for CO, NO_x, SO₂, O₂ and CO₂

Manufacturer: Horiba GmbH
Kaplanstrasse 5
A-3430 Tulln
Austria

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**


**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


**15 Notification as regards Federal Environment Agency (UBA) notices
of 23 February 2012 (BAnz. p. 920, chapter I number 4.5) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 41)**

The ENDA 5000 measuring device with CMA-5800 E analyzer module for NO_x, SO₂, CO, CO₂ and O₂ from HORIBA GmbH can also be equipped with the KPMW-MT/TC102 heating element from Kashima Co., Ltd. for heating the pump housing.

Statement issued by TÜV Rheinland Energy GmbH dated 28 April 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.