

# CONFIRMATION

Notification: 0000037051\_02\_01  
on changes according to EN 15267 regarding certificate 0000037051\_02 dated 20 July 2022

---

**AMS designation:** GM32 In-Situ-Gasanalysator Cross Duct for NO and SO<sub>2</sub>

**Manufacturer:** SICK AG  
Gisela Sick Straße 1  
79276 Reute  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 28 June 2022  
Publication: BAnz AT 28.07.2022 B4**

**III. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**


**55 Notification as regards Federal Environment Agency (UBA) notices  
of 3 August 2009 (BAnz. p. 2929, chapter I number 3.2) and  
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter VI correction 1)**


The GM32 Cross Duct measuring device for NO and SO<sub>2</sub> from SICK AG is also suitable  
for use at plants complying with the 44th BImSchV.

Statement issued by TÜV Rheinland Energy GmbH dated 12 May 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 01 August 2022

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.