

CONFIRMATION

Notification: 0000038505_04_01
on changes according to EN 15267 regarding certificate 0000038505_04 dated 02 March 2023

AMS designation: LasIR HF for HF

Manufacturer: Unisearch Associates
96 Bradwick Drive
Concord, Ontario / L4K 1K8
Canada

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 05 July 2023
Publication: BAnz AT 02.08.2023 B7**

**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


- 26 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter I number 2.1) and
of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 32)**


The control/analysis unit of the LasIR measuring system for HF from Unisearch Associates Inc. may alternatively be installed in a housing with a depth of 12" instead of 10.5".

The measuring system can alternatively be equipped with the enlarged LCD screen PCA-LCD2000.

Statement issued by TÜV Rheinland Energy GmbH dated 02 February 2023

TÜV Rheinland Energy GmbH
Cologne, 02 August 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.