

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040205_02

AMS designation: Serinus 50 for SO₂

Manufacturer: Ecotech Pty Ltd.
1492 Ferntree Gully Road
Knoxfield, VIC, 3180
Australia

Test Laboratory: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards:
VDI 4202-1 (2010), VDI 4203-3 (2010), EN 14212 (2012),
EN 15267-1 (2009) and EN 15267-2 (2009).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 13 pages).
The present certificate replaces certificate 0000040205_01 of 01 April 2019.



Suitability Tested
Equivalent to
2008/50/EC
EN 15267
Regular Surveillance
www.tuv.com
ID 0000040205

Publication in the German Federal Gazette
(BAnz) of 01 April 2014


This certificate will expire on:
30 June 2025

German Federal Environment Agency
Dessau, 01 July 2020

TÜV Rheinland Energy GmbH
Cologne, 30 June 2020



Dr. Marcel Langner
Head of Section II 4.1



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone: + 49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to certificate D-PL-11120-02-00.

Test Report:	936/21221977/B dated 08 October 2013
Initial certification:	01 April 2014
Expiry date:	30 June 2025
Certificate:	Renewal (of previous certificate 0000040205_01 dated 01 April 2019 valid until 30 June 2020)
Publication:	BAnz AT 01.04.2014 B12, chapter IV number 3.1

Approved application

The certified AMS is suitable for continuous ambient air monitoring of sulphur dioxide (stationary operation).

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-months field test.

The AMS is approved for an ambient temperature range of 0 °C to +30 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for monitoring the AMS readings relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended purpose.

Basis of the certification

This certification is based on:

- Test report no. 936/21221977/B dated 08 October 2013 issued by TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, chapter IV number 3.1,
UBA announcement dated 27 February 2014:

AMS designation:

Serinus 50 for SO₂

Manufacturer:

Ecotech Pty Ltd., Knoxfield, Australia

Field of application:

Continuous measurement of sulphur dioxide concentration in ambient air (stationary operation)

Measuring range during performance testing:

Component	Certification range	Unit
Sulphur dioxide	0–1 000	µg/m ³

Software version:

Firmware: 2.09.0005

Restrictions:

None

Notes:

1. The measuring system must be operated inside a lockable measuring cabinet or measurement container.
2. The test report on performance testing is available on the internet at www.qal1.de.

Test Laboratory:

TÜV Rheinland Energie und Umwelt GmbH, Cologne
Report no.: 936/21221977/B dated 08 October 2013

Publication in the German Federal Gazette: BAnz AT 02.04.2015 B5, chapter IV notification 7, UBA announcement dated 25 February 2015:

7 Notification as regards Federal Environment Agency (UBA) notice of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1).

The Serinus 50 measuring system for SO₂ manufactured by Ecotech Pty Ltd. will be equipped with a new micro processor board (C010014) in the future. This entails changes to the power supply and the software.

The following software versions apply:

2.20.0009 for instruments with the previous processor board (C010001)
3.10.001 for instruments with the new processor board (C010014)

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 12 September 2014

Publication in the German Federal Gazette: BAnz AT 15.03.2017 B6, chapter V notification 8, UBA announcement dated 22 February 2017:

8 Notification as regards Federal Environment Agency (UBA) notices of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter IV 7th notification)

The latest software version of the Serinus 50 measuring system for SO₂ with microprocessor C010001 manufactured by Ecotech Pty Ltd. is:
V 2.31.0004.

Moreover, the following software version are approved for this instrument version:
V 2.21.0000, V 2.22.0000, V 2.23.0000, V 2.24.0000, V 2.25.0004, V 2.26.0000,
V 2.27.0000, V 2.28.0000, V 2.29.0003 und V 2.30.0000.

The latest software version of the Serinus 50 measuring system for SO₂ with microprocessor C010014 manufactured by Ecotech Pty Ltd. is:
V 3.48.011.

Moreover, the following software version are approved for this instrument version:
V 3.13.000, V 3.14.001, V 3.15.010, V 3.16.001, V 3.18.003, V 3.20.000,
V 3.22.000, V 3.23.015, V 3.24.000, V 3.26.000, V 3.27.000, V 3.28.000,
V 3.29.013, V 3.30.005, V 3.31.002, V 3.32.003, V 3.33.004, V 3.34.000,
V 3.35.004, V 3.36.000, V 3.37.004, V 3.38.006, V 3.39.000, V 3.40.001,
V 3.41.004, V 3.42.000, V 3.43.000, V 3.44.004, V 3.45.011, V 3.46.002,
V 3.47.006.

Statement issued by TÜV Rheinland Energy GmbH dated 13 October 2016

Publication in the German Federal Gazette: BAnz AT 26.03.2019 B7, chapter IV notification 18, UBA announcement dated 27 February 2019:

18 Notification as regards Federal Environment Agency (UBA) notices of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter IV 8th notification)

The latest software version of the Serinus 50 measuring system for SO₂ with microprocessor C010001 manufactured by Ecotech Pty Ltd. is:
V 2.35.0001.

Moreover, the following software version are approved for this instrument version:
V 2.32.0000, V 2.33.0000, V 2.34.0000

The latest software version of the Serinus 50 measuring system for SO₂ with microprocessor C010014 manufactured by Ecotech Pty Ltd. is:
V 3.74.0003.

Moreover, the following software version are approved for this instrument version:
V 3.49.0000, V 3.51.0011, V3.52.0000, V 3.53.0012, V 3.54.0000, V 3.55.0000,
V 3.56.0001, V 3.57.0002, V 3.58.0000, V 3.59.0004, V 3.60.0005, V 3.61.0000,
V 3.62.0000, V 3.63.0001, V 3.64.0000, V 3.65.0001, V 3.66.0000, V 3.67.0003,
V 3.68.0009, V 3.69.0001, V 3.70.0000, V 3.71.0000

The instrument's display shows the software version in the following format:
2.XX or 3.XX.

Statement issued by TÜV Rheinland Energy GmbH dated 10 October 2018

Publication in the German Federal Gazette: BAnz AT 24.03.2020 B7, chapter IV notification 22, UBA announcement dated 24 February 2020:

22 Notification as regards Federal Environment Agency (UBA) notices of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter IV 18th notification)

The latest software version of the Serinus 50 measuring system for SO₂ with microprocessor C010001 manufactured by Ecotech Pty Ltd. remains:
V 2.35.0001.

The latest software version of the Serinus 50 measuring system for S_{O2} with microprocessor C010014 manufactured by Ecotech Pty Ltd. is:
V 3.87.0000.

Moreover, the following software version are approved for this instrument version:
V 3.75.0003, V 3.76.0004, V 3.77.0009, V 3.78.0000, V 3.79.0001, V 3.81.0000,
V 3.83.0000, V 3.84.0000, V 3.85.0001, V 3.86.0000.

The instrument's display shows the software version in the following format:
2.XX or 3.XX.

Statement issued by TÜV Rheinland Energy GmbH dated 20 September 2019

Certified product

This certification applies to automated measurement systems conforming to the following description:

The Serinus 50 measuring system is a continuous sulphur dioxide monitor which uses the method of ultraviolet fluorescence. The instrument is designed for the continuous measuring of sulphur dioxide in ambient air.

Measurements are performed by means of the following components:

- Hydrocarbon kicker
- UV lamp
- fluorescence cell
- optical band-pass filter
- photomultiplier tube (PMT)

The SO₂ concentration is automatically corrected for gas temperature and pressure changes and referenced to 0 °C, 20 °C or 25 °C at 1 atmosphere. This allows the Serinus 50 to sample in the most useful range of SO₂ ambient measurement (25–500 ppb SO₂ in the air).

The measurement of sulphur dioxide is based on classical fluorescence spectroscopy principles. Sulphur dioxide (SO₂) exhibits a strong ultraviolet (UV) absorption spectrum between 200 and 240 nm. When SO₂ absorbs UV from this wavelength, photon emission occurs (300–420 nm). The amount of fluorescence emitted is directly proportional to the SO₂ concentration.

The Serinus 50 follows these principles and measurement techniques:

- Sample air passes through a hydrocarbon kicker which removes hydrocarbons.
- UV energy from the zinc discharge lamp passes through a UV band-pass filter are used to produce radiation at 214 nm.
- The radiation is focused into the fluorescence cell where it is absorbed by the SO₂ molecules.
- The SO₂ molecules then emit photons (fluorescent light) uniformly in all directions.
- Wavelengths between 310–350 nm, which are specific to SO₂, pass through a band pass filter where they reach the photomultiplier and record a signal. The signal is recorded accordingly.
- A reference detector monitors the emission from the zinc lamp and is used to correct for fluctuations in lamp intensity.

Exhaust air is scrubbed with a charcoal scrubber to eliminate hydrocarbons and SO₂. This air is then clean enough for use in the hydrocarbon kicker to remove hydrocarbons from the incoming sample air.

The Serinus 50 sulphur dioxide analyser consists of five main assemblies:

- The pneumatics to transfer sample and exhaust gas,
- The sensors for the measurement of SO₂ (optical cell) and other relevant parameters,
- The control system which encompasses all circuit boards controlling sensors and pneumatic,
- The power supply which supplies power for all the instrument processors,
- The communication module to access data.

Particle filter:

The particulate filter is a Teflon 5 micron (µm) filter with a diameter of 47 mm. This filter eliminates all particles larger than 5 µm that could interfere with sample measurements.

Hydrocarbon kicker

The hydrocarbon kicker removes interfering hydrocarbons from the sample air. To this effect a counter current exchange is used, where an air with a lower concentration of hydrocarbons moves in an opposite direction to air with a higher concentration. The high concentrations of hydrocarbons diffuse through a selective permeation membrane to the low concentration exhaust air and are removed. Increasing the flow of the low concentration air also increases the rate of diffusion.

Sample gas pump

Manufacturer: Thomas, Type: 617CD22-194 C

During performance testing, the sample gas pump mentioned above was used for the laboratory as well as in the field test. As far as the models Serinus 10 (ozone), Serinus 30 (CO) and Serinus 50 (SO₂) are concerned, one pump can be operated with up to two analysers. However, operation of the Serinus 40 (NO_x) requires one sample gas pump per analyser.

General remarks

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacturing process for the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. Upon revocation of the publication the certificate loses its validity. After the expiration of the certificate and on request of TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must no longer be used.

The relevant version of this certificate and its expiration date are also accessible on the internet at qal1.de.

Document history

Certification of the Serinus 50 measuring system is based on the documents listed below and the regular, continuous surveillance of the manufacturer's quality management system:

Initial certification according to EN 15267

Certificate no. 0000040205: 29 April 2014
Expiry date of the certificate: 31 March 2019
Test report no.: 936/21221977/B dated 8 October 2013
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAnz AT 01.04.2014 B12, chapter IV number 3.1
UBA announcement dated 27 February 2014

Notifications in accordance with EN 15267

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 12 September 2014
Publication: BAnz AT 02.04.2015 B5, chapter IV notification 7
UBA announcement dated 25 February 2015
(Design and software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 13 October 2016
Publication: BAnz AT 15.03.2017 B6, chapter V notification 8
UBA announcement dated 22 February 2017
(software updates)

Renewal of the certificate

Certificate no. 0000040205_01: 01 April 2019
Expiry date of the certificate: 30 June 2020

Notifications in accordance with EN 15267

Statement issued by TÜV Rheinland Energy GmbH dated 10 October 2018
Publication: BAnz AT 26.03.2019 B7, chapter IV notification 18
UBA announcement dated 27 February 2019
(software updates)

Statement issued by TÜV Rheinland Energy GmbH dated 20 September 2019
Publication: BAnz AT 24.03.2020 B7, chapter IV notification 22
UBA announcement dated 24 February 2020
(software updates)

Renewal of the certificate

Certificate no. 0000040205_02: 01 July 2020
Expiry date of the certificate: 30 June 2025

Expanded uncertainty from the results obtained in the laboratory tests for analyser 1

Measuring device:		Serial No.:		13-0086 (Device 1)	
Measured component:		1h-limit value:		132 mmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.050	u_z	0.0002
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.240	$u_{r,h}$	0.0050
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	2.620	$u_{l,h}$	3.9868
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.340	u_{gp}	7.2852
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.050	u_{gt}	0.1609
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.305	u_{st}	6.1146
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.027	u_v	0.0608
8a	Interferent H ₂ O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	0.010	u_{p20}	5.0688
8b	Interferent H ₂ S with 200 nmol/mol	≤ 10 nmol/mol (Span)	3.040	$u_{r,1,po5}$	
8c	Interferent NH ₃ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	1.600		
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Span)	2.390		
8e	Interferent NO ₂ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.290		
8f	Interferent m-Xylene with 1 µmol/mol	≤ 5.0 nmol/mol (Span)	1.080		
9	Averaging effect	≤ 5.0 nmol/mol (Zero)	3.420		
18	Difference sample/calibration port	≤ 5.0 nmol/mol (Span)	2.850		
21	Uncertainty of test gas	≤ 5.0 nmol/mol (Zero)	0.100		
		≤ 10 nmol/mol (Span)	0.740		
		≤ 7.0% of measured value	3.050		
		≤ 1.0%	-2.930	$u_{r1,ng}$	
		≤ 3.0%	0.220	$u_{b,v}$	4.9861
			2.000	$u_{b,c}$	0.0843
				u_{gg}	1.7424
				u_c	7.9689
				U	15.9379
				W	12.07
				W_{req}	15

Expanded uncertainty from the results obtained in the laboratory tests for analyser 2

Measuring device: Ecotech Serinus 50		Serial-No.: 13-0097 (Device 2)	132		nmol/mol
Measured component: SO ₂		1h-limit value:			
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.000	U _{r,z} 0.00	0.0000
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.230	U _{r,h} 0.07	0.0048
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	1.590	U _{l,h} 1.21	1.4683
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.270	U _{sp} 2.14	4.5625
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.030	U _{gt} 0.24	0.0587
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.152	U _{st} 1.24	1.5295
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.028	U _v 0.26	0.0701
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	-0.510	U _{h₂O} 2.11	4.4660
		≤ 10 nmol/mol (Span)	3.060		
8b	Interferent H ₂ S with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	1.410	U _{h₂S,sp}	
		≤ 5.0 nmol/mol (Span)	2.210		
8c	Interferent NH ₃ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.310		
		≤ 5.0 nmol/mol (Span)	0.230		
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Zero)	3.670	5.48	30.0628
		≤ 5.0 nmol/mol (Span)	4.160	or	
8e	Interferent NO ₂ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	1.000		
		≤ 5.0 nmol/mol (Span)	0.310		
8f	Interferent m-Xylene with 1 µmol/mol	≤ 10 nmol/mol (Zero)	0.860	U _{h₂,neg}	
		≤ 10 nmol/mol (Span)	2.660		
9	Averaging effect	≤ 7.0% of measured value	-2.620	U _{av} -2.00	3.9868
18	Difference sample/calibration port	≤ 1.0%	0.280	U _{diff,c} 0.37	0.1366
21	Uncertainty of test gas	≤ 3.0%	2.000	U _{tg} 1.32	1.7424
Combined standard uncertainty			u _c		6.9346
Expanded uncertainty			U		13.8692
Relative expanded uncertainty			W		10.51
Maximum allowed expanded uncertainty			W _{req}		15

Expanded uncertainty from the results obtained in the laboratory and field tests for analyser 1

Measuring device: Ecotec h Serimus 50		Serial-No.: 13-0096 (Device 1)		132		nmol/mol	
Measured component: SO ₂		1h-limit value:		Partial uncertainty		Square of partial uncertainty	
No.	Performance characteristic	Performance criterion	Result	U _z	U _{i,h}	U _{i,h}	U _{i,h}
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.050	U _z	0.02	0.0002	
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.240	U _{i,h}	not considered, as ur.lh = 0.07 < ur.f	-	
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	2.620	U _{i,h}	2.00	3.9868	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.340	U _{ip}	2.70	7.2852	
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.050	U _{gt}	0.40	0.1609	
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.305	U _{st}	2.47	6.1146	
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.027	U _v	0.25	0.0608	
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.010				
8b	Interferent H ₂ S with 200 nmol/mol	≤ 10 nmol/mol (Span)	3.040	U _{H2O}	2.25	5.0688	
8c	Interferent NH ₃ with 200 nmol/mol	≤ 5.0 nmol/mol (Span)	1.600	U _{interpos}			
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Zero)	2.390				
8e	Interferent NO ₂ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.290				
8f	Interferent m-Xylene with 1 µmol/mol	≤ 5.0 nmol/mol (Span)	1.080				
9	Averaging effect	≤ 5.0 nmol/mol (Zero)	3.420		5.83	34.0086	
10	Reproducibility standard deviation under field conditions	≤ 5.0 nmol/mol (Span)	2.850	U _{int,neg}			
11	Long term drift at zero level	≤ 7.0% of measured value	0.100	U _{av}	-2.23	4.9861	
12	Long term drift at span level	≤ 5.0% of average over 3 months	0.740	U _{av}	4.94	24.3720	
18	Difference sample/calibration port	≤ 5.0% of max. of certification range	-0.940	U _{i,f}	-0.54	0.2945	
21	Uncertainty of test gas	≤ 1.0%	3.810	U _{li,h}	2.90	8.4310	
		≤ 3.0%	0.220	U _{ass}	0.29	0.0843	
			2.000	U _g	1.32	1.7424	
		Combined standard uncertainty		U _c		9.8283	nmol/mol
		Expanded uncertainty		U		19.6567	nmol/mol
		Relative expanded uncertainty		W		14.89	%
		Maximum allowed expanded uncertainty		W _{req}		15	%

Expanded uncertainty from the results obtained in the laboratory and field tests for analyser 2

Measuring device: Ecotech Serinus 50		Serial-No.: 13-0097 (Device 2)	1h-limit value: 132		nmol/mol
Measured component: SO ₂		Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
No.	Performance characteristic	≤	0.000	U _{i,z}	0.0000
1	Repeatability standard deviation at zero	≤	0.000	U _{i,z}	0.0000
2	Repeatability standard deviation at 1h-limit value	≤	0.230	U _{i,th}	not considered, as U _{i,th} = 0.06 < U _{r,f}
3	"lack of fit" at 1h-limit value	≤	1.590	U _{i,th}	1.4683
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤	0.270	U _{i,sp}	4.5625
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤	0.030	U _{i,t}	0.0587
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤	0.152	U _{i,st}	1.5295
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤	0.028	U _i	0.0701
8a	Interferent H ₂ O with 21 nmol/mol	≤	-0.510		
8b	Interferent H ₂ S with 200 nmol/mol	≤	3.060	U _{i,po}	4.4660
8c	Interferent NH ₃ with 200 nmol/mol	≤	1.410	U _{i,pos}	
8d	Interferent NO with 500 nmol/mol	≤	2.210		
8e	Interferent NO ₂ with 200 nmol/mol	≤	-0.310		
8f	Interferent m-Xylene with 1 µmol/mol	≤	0.230		
9	Averaging effect	≤	3.670	or	30.0628
10	Reproducibility standard deviation under field conditions	≤	1.000		
11	Long term drift at zero level	≤	0.310		
12	Long term drift at span level	≤	0.860		
18	Difference sample/calibration port	≤	2.660	U _{i,reg}	
21	Uncertainty of test gas	≤	-2.620	U _{av}	3.9868
		≤	3.740	U _f	24.3720
		≤	1.470	U _{cl,z}	0.7203
		≤	3.540	U _{j,th}	7.2784
		≤	0.280	U _{asc}	0.1366
		≤	2.000	U _g	1.7424
				U _e	8.9696
				U	17.9393
				W	13.59
				W _{reg}	15
					%