

CONFIRMATION

Notification: 0000040212_05_01
on changes according to EN 15267 regarding certificate 0000040212_05 of 12 June 2019

AMS designation: Fidas® 200 S, Fidas® 200 E and Fidas® 200 for PM₁₀ and PM_{2,5}

Manufacturer: PALAS GmbH
Greschbachstraße 3b
76229 Karlsruhe
Germany

German Federal Environment Agency (UBA)

**Announcement about the uniform practice in
monitoring emissions and ambient air
dated 28 June 2019
Publication: BAnz AT 22.07.2019 B8**

V. Notification as regards the Uniform Practice in monitoring emissions:

**14 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and
of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter IV 44th notification)**

The software version of the Fidas® 200 S, Fidas® 200 E and Fidas® 200 measuring system for PM₁₀ and PM_{2,5} has been updated. The latest software version is

100451.0014.0001.0001.0011.


Alongside this version, the following intermediary versions can also be used:

100450.0014.0001.0001.0011.


At present, it is possible to use the Fidas sensor with or without insect protection. In future, an insect protection ring with additional seal and cement for the remaining slit will be available. The measuring system can also be operated with the Lufft WS500-UMB weather station.

Statement issued by TÜV Rheinland Energy GmbH dated 6 March 2019

TÜV Rheinland Energy GmbH
Cologne, 15 September 2019



i. V. Dipl.-Ing. Guido Baum



i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.