

CONFIRMATION

Notification: 0000040212_05_03
on changes according to EN 15267 regarding certificate 0000040212_05 dated 12 June 2019

AMS designation: Fidas 200 S Fidas 200 Fidas 200 E for PM₁₀, PM_{2.5}

Manufacturer: PALAS GmbH
Greschbachstrasse 3b
76229 Karlsruhe
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 27 May 2020**

Publication: Federal Gazette: BAnz AT 31.07.2020 B10

**II. Notifications as regards the uniform practice in monitoring
emissions and ambient air**

**16 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and
of 31 March 2020 (BAnz AT 07.05.2020 B8, chapter III, notification 4)**


The software version of the Fidas® 200 S, Fidas® 200 E and Fidas® 200 measuring system for PM₁₀ and PM_{2.5} has been updated. The current software version is: 100465.0014.0001.0001.0011


In addition to this version, the following intermediate versions are also valid: 100464.0014.0001.0001.0011

The Fidas® 200 S, Fidas® 200E or Fidas® 200 measuring system may alternatively be equipped with the Hamamatsu H10721-210 photomultiplier and with the panel PC Pico318-N3350 manufactured by Axiomtek incl. touch screen DLC 0700 manufactured by DLC.

Statement issued by TÜV Rheinland Energy GmbH dated 07 May 2020

TÜV Rheinland Energy GmbH
Cologne, 07 August 2020


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.