

CONFIRMATION

Notification: 0000040212_05_05
on changes according to EN 15267 regarding certificate 0000040212_05 dated 12 June 2019

AMS designation: Fidas 200 S, Fidas 200, Fidas 200 E for PM₁₀ and PM_{2.5}

Manufacturer: PALAS GmbH
Greschbachstrasse 3b
76229 Karlsruhe
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 09 March 2022
Publication: BAnz AT 11.04.2022 B10**

**VI. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**21 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 45)**

The measuring devices Fidas® 200 S, Fidas® 200 E or Fidas® 200 PM₁₀ and PM_{2.5} of the company Palas GmbH can be equipped in the future alternatively with the LED SBM-40-SC of the manufacturer Luminus. As an alternative to the previously used sampling pump of type 1420VDP BLDC from the company Thomas Pumps, the 2-head diaphragm pump of type NMP830.1.2KPDC-B HP 24V from the company KNF can be used in the future.


In the future, the sampling tube can be varied in length between 1.2 and 2 m according to customer requirements.


The current software version is: 100525.0014.0001.0001.0011

In addition to this version number, the following intermediate versions are also valid:
100468.0014.0001.0011 and 100524.0014.0001.0001.0011

Statement issued by TÜV Rheinland Energy GmbH dated 08 December 2021

TÜV Rheinland Energy GmbH
Cologne, 12 April 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.