

CONFIRMATION

Notification: 0000040212_05_06
on changes according to EN 15267 regarding certificate 0000040212_05 dated 12 June 2019

AMS designation: Fidas 200 S, Fidas 200, Fidas 200 E for PM₁₀, PM_{2,5}

Manufacturer: PALAS GmbH
Greschbachstrasse 3b
76229 Karlsruhe
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**

**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**44 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and
of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 21)**

The measuring devices Fidas® 200 S, Fidas® 200 E resp. Fidas® 200 for PM₁₀ and PM_{2,5} of the company Palas GmbH can be equipped alternatively with the weather station HTP-Geber compact with active ventilated weather protection of the company Thies from the software version 100532.0014.0001.0011 on.

For measuring the LED temperature, the sensor B57861S0103F040 from TDK can be used as an alternative in the future.


For the connection of the weather station and the IADS module the plugs WSV 50 or SV 50 (weather station) and WSV 60 or SV 60 (IADS module) of the company Lumberg can be used alternatively.


The current software version is: 100532.0014.0001.0001.0011

In addition to this version number, the following intermediate version is also valid:
100527.0014.0001.0011.

Statement issued by TÜV Rheinland Energy GmbH dated 28 April 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.