

CONFIRMATION

Notification: 0000040212_05_08
on changes according to EN 15267 regarding certificate 0000040212_05 dated 12 June 2019

AMS designation: Fidas 200 S, Fidas 200, Fidas 200 E for PM₁₀ and PM_{2,5}

Manufacturer: PALAS GmbH
Greschbachstrasse 3b
76229 Karlsruhe
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 05 July 2023
Publication: BAnz AT 02.08.2023 B7**

**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**36 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 80)**

The current software version for the Fidas® 200 S, Fidas® 200 E / Fidas® 200
measuring systems for PM₁₀ and PM_{2,5} from Palas GmbH is:

100537.0014.0001.0001.0011


In addition to this version number, the following intermediate version is also valid:


100536.0014.0001.0001.0011

If the PC Pico318-N3350 is installed, it can also have a memory capacity of 128 GB
instead of 32 GB up to now.

Statement issued by TÜV Rheinland Energy GmbH dated 31 March 2023

TÜV Rheinland Energy GmbH
Cologne, 02 August 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.