

CONFIRMATION

Notification: 0000040213_02_01
on changes according to EN 15267 regarding certificate 0000040213_02 dated 01 July 2020

AMS designation: Modell 5014i Beta for PM_{2,5}

Manufacturer: Thermo Fisher Scientific
27, Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 29 June 2021
Publication: BAnz AT 05.08.2021 B5**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

- 4 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 6.2) and
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter V notification 20)**


The measuring head of the ambient air monitoring system, Model 5014i Beta, with PM_{2,5} pre-separator for PM_{2,5} suspended particulate matter manufactured by the company Thermo Fisher Scientific was modified to increase operational safety with regard to potential leakages.


The latest software version of the measuring system is:
V 03.00.01 (111578-00)

In addition to this revision number, the following interim version is also valid:
V 03.00.00 (111578-00)

Statement issued by TÜV Rheinland Energy GmbH dated 24 February 2021

TÜV Rheinland Energy GmbH
Cologne, 16 August 2021


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.