

CONFIRMATION

Notification: 0000040218_02_02
on changes according to EN 15267 regarding certificate 0000040218_02 dated 01 July 2020

AMS designation: Modell 43i for SO₂

Manufacturer: Thermo Fisher Scientific
27, Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**

III. Notifications as regards the uniform practice in monitoring emissions and ambient air:


46 Notification as regards Federal Environment Agency (UBA) notices of 12 September 2006 (BAnz. p. 6715, chapter IV number 2.2) and of 29 June 2021 (BAnz AT 05.08.2021 B5, chapter IV notification 10)


The following hardware changes have been introduced for the Model 43i measuring device for SO₂ from Thermo Fisher Scientific:

- The measuring device can now also be equipped with the Arcturus CPU (53281) processor board.
- The measuring device can now also be equipped with a SUNON housing fan (Art. No. PMD2408PMB-A).

Statement issued by TÜV Rheinland Energy GmbH dated 18 May 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.