

CONFIRMATION

Notification: 0000043107_02_01
on changes according to EN 15267 regarding certificate 0000043107_02 of 12 June 2019

AMS designation: APDA-372 and APDA-372E for PM₁₀ and PM_{2,5}
Manufacturer: HORIBA Europe GmbH
Hans-Mess-Str. 6
61440 Oberursel/Ts.
Germany

German Federal Environment Agency (UBA)

**Announcement about the uniform practice in
monitoring emissions and ambient air
dated 28 June 2019
Publication: BAnz AT 22.07.2019 B8**

V. Notification as regards the Uniform Practice in monitoring emissions:

13 Notification as regards Federal Environment Agency (UBA) notices of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 3.1) and of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter IV 38th notification)

The software of the APDA-372 and APDA-372E for PM₁₀ and PM_{2,5} manufactured by HORIBA Europe GmbH has been updated. The current software version is:

100451.0014.0001.0001.0011.

Alongside this version, the following intermediary versions can also be used:


100450.0014.0001.0001.0011.


At present, it is possible to use the APDA-372 sensor with or without insect protection. In future, an insect protection ring with additional seal and cement for the remaining slit will be available.

In future, the length of the sampling tube may vary between 1.2 m and 2 m according to customers' specifications. The measuring system can also be operated with the Lufft WS500-UMB weather station.

Statement issued by TÜV Rheinland Energy GmbH dated 6 March 2019

TÜV Rheinland Energy GmbH
Cologne, 15 September 2019


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to certificate D-PL-11120-02-00.