

CONFIRMATION

Notification: 0000043107_02_06
on changes according to EN 15267 regarding certificate 0000043107_02 dated 12 June 2019

AMS designation: APDA-372 / APDA-372 E for PM₁₀ and PM_{2.5}

Manufacturer: HORIBA Europe GmbH
Hans-Mess-Str. 6
61440 Oberursel /Ts.
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**77 Notification as regards Federal Environment Agency (UBA) notices
of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 3.1) and
of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 34)**

The current software version of the APDA-372 or APDA-372 E measuring systems for
PM₁₀ and PM_{2.5} from the company HORIBA Europe GmbH is:


100535.0014.0001.0001.0011


The operational amplifiers on the SLA board can be installed either as THT components
or as SMD components.

The measuring system can alternatively be operated with the weather station EVA730
from Kroneis.

Statement issued by TÜV Rheinland Energy GmbH dated 16 September 2022

TÜV Rheinland Energy GmbH
Cologne, 25 April 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.