

CONFIRMATION

Notification: 0000043107_02_07
on changes according to EN 15267 regarding certificate 0000043107_02 dated 12 June 2019

AMS designation: APDA-372 / APDA-372 E for PM₁₀ and PM_{2,5}

Manufacturer: HORIBA Europe GmbH
Hans-Mess-Str. 6
61440 Oberursel /Ts.
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 05 July 2023
Publication: BAnz AT 02.08.2023 B7**

III. Notifications as regards the uniform practice in monitoring emissions and ambient air:

35 Notification as regards Federal Environment Agency (UBA) notices of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 3.1) and of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 77)

The current software version for the APDA-372 / APDA-372 E measuring systems for
PM₁₀ and PM_{2,5} from HORIBA Europe GmbH is:

100537.0014.0001.0001.0011


In addition to this version number, the following intermediate version is also valid:


100536.0014.0001.0001.0011

If the PC Pico318-N3350 is installed, it can also have a memory capacity of 128 GB
instead of 32 GB up to now.

Statement issued by TÜV Rheinland Energy GmbH dated 31 March 2023

TÜV Rheinland Energy GmbH
Cologne, 02 August 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.