

# CONFIRMATION

Notification: 0000053810\_12\_01  
on changes according to EN 15267 regarding certificate 0000053810\_12 dated 03 September 2021

---

**AMS designation:** Set CEM CERT 7MB1957 for CO, NO, NO<sub>2</sub>, NO<sub>x</sub>, SO<sub>2</sub>, O<sub>2</sub> and CO<sub>2</sub>

**Manufacturer:** Siemens  
Östliche Rheinbrückenstr. 50  
76187 Karlsruhe  
Germany

---

## German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 28 June 2022  
Publication: BAnz AT 28.07.2022 B4**

### III. Notifications as regards the uniform practice in monitoring emissions and ambient air:

#### 29 Notification as regards Federal Environment Agency (UBA) notice of 29 June 2021 (BAnz AT 05.08.2021 B5, chapter I number 4.2)

The current software versions of the measuring device Set CEM CERT 7MB1957 for  
CO, NO, NO<sub>2</sub>, SO<sub>2</sub>, NO<sub>x</sub>, CO<sub>2</sub> and O<sub>2</sub> from Siemens AG are:

Ultramat 23-7MB2355 4.02.09, Ultramat 23-7MB2357 4.02.09

Ultramat 23-7MB2358 4.02.09, Ultramat 6 4.8.8, Ultramat 6-2K 4.8.8, Oxymat 6 4.8.8

Ultramat/Oxymat 6 4.8.8, SIEMENS SIMATIC Set CEM CERT 7MB1957 Rev. 2.0

SIPROCESS UV600-7MB2621, BCU 9150883\_3.003

Gas module 9137582\_3.002, UV module 9139736\_3.005

SIPROCESS GA700 Ultramat 7 CALC 1.70.00 / ADU 1.40.02

SIPROCESS GA700 Oxymat 7 CALC 1.40.08 / ADU 1.30.00


A new, equivalent chopper motor is used in the measuring modules of the Ultramat 6  
series.


An alternative manufacturer has been qualified for the bandwidth filter used in the IR  
measuring modules of the Ultramat 23 and Ultramat 6 series for the determination of NO  
and SO<sub>2</sub>.

Statement issued by TÜV Rheinland Energy GmbH dated 09 February 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 01 August 2022

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.