

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000059869

AMS designation: HM-1400 TRX 2 for Hg

Manufacturer: DURAG GmbH
Kollaustraße 105
22453 Hamburg
Germany

Test Laboratory: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested and certified
according to the standards**

**EN 15267-1: 2009, EN 15267-2: 2009, EN 15267-3: 2007
and EN 14181: 2014.**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 6 pages).



Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance


www.tuv.com
ID 0000059869

Publication in the German Federal Gazette
(BAnz) of 17 July 2018


This certificate will expire on:
16 July 2023

German Federal Environment Agency
Dessau, 4 September 2018

TÜV Rheinland Energy GmbH
Cologne, 3 September 2018



Dr Marcel Langner
Head of Section II 4.1



ppa. Dr Peter Wilbring

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone: + 49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

Test Report:	936/21238805/C dated 10 May 2018
Initial certification:	17 July 2018
Expiry date:	16 July 2023
Publication:	BAnz AT 17.07.2018 B9, chapter I number 2.1

Approved application

The tested AMS is suitable for use at combustion plants according to Directive 2010/75/EU, chapter III (13th BImSchV). The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a four-months field test at a large combustion plant (hard coal firing).

The AMS is approved for an ambient temperature range of +5 °C to +40 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the limit values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Basis of the certification

This certification is based on:

- Test report 936/21238805/C dated 10 May 2018 issued by TÜV Rheinland Energy GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 17.07.2018 B9, chapter I number 2.1,
UBA announcement dated 3 July 2018:

AMS designation:

HM-1400 TRX 2 for mercury

Manufacturer:

DURAG GmbH, Hamburg

Field of application:

For plants according to the 13th BImSchV

Measuring ranges during performance testing:

Component	Certification range	supplementary measuring ranges		Unit
Hg	0–15	0–45	0–75	µg/m ³

Software versions:

SPS: 3.01R000
Display: TRX_3.01R0000

Restrictions:

None

Notes:

1. The maintenance interval is four weeks.
2. Wet test gases should be used for testing mercury.
3. An external test gas generator is needed for regular span checks during the maintenance interval.
4. The probe length was 40 m during performance testing.
5. The zero point is automatically adjusted every two hours using purified ambient air.
6. Manual QAL3 tests and automatic span point checks should not be performed the same day.

Test Report:

TÜV Rheinland Energy GmbH, Cologne
Report no.: 936/21238805/C dated 10 May 2018

Certified product

This certification applies to automated measurement systems conforming to the following description:

The extractive HM-1400 TRX 2 measuring system is an AMS for the continuous analysis of all gaseous mercury compounds emitted by plants subject to monitoring (currently Directive 2010/75/EU, chapter III \triangleq 13th BImSchV) or present in process gases. A dual beam photometer detects the concentration of atomic mercury Hg^0 . In order to determine total mercury Hg (total) in the sample gas, mercury present in the sample gas is first reduced to Hg^0 . This takes place in a thermo-catalytic reactor.

The AMS provides the following control functions:

- Leak test: The AMS tests the leak tightness of the system.
- Zero point checks: The AMS automatically performs zero point checks for an internal re-adjustment of the photometer.
- Span point measurement: With its internal calibration gas generator (AKM), the instrument performs an automatic span point measurement. This feature has not yet been QAL3-approved for regular checks. Thus, the use of the internal AKM does not conform to the certified version of the measuring system.
- Connecting external gas generators: This facilitates to connect external gas generators for checking the photometer and the entire system. In the context of QAL3-compliant performance, this is typically done via an appropriate stub at the sampling probe for dosing test gas upstream of the filter.

The measuring system provides options for sample gas dilution and separate determination of the proportions of Hg species Hg^0 and $\text{Hg}^{\text{n}+}$ (specification). As soon as this option is activated, the measuring system does no longer conform to the certified version. Measured values obtained in this operation mode must not be used for continuous emission monitoring.

The essential components of the HM-1400 TRX 2 measuring system include:

- Measuring cabinet c/w photometer (UV dual beam photometer), thermo-catalytic reactor (with two reactor vessels which can be switched automatically), pressure and flow regulator, temperature control and test gas generator
- M&C sample probe SP2000-H with two stubs upstream of the filter
- Heated sampling line (185 °C) with two internal lines (6 mm PTFE, one for sucking the sample gas downstream of the filter and one for dosing in zero and test gases upstream of the filter). The sampling line used for the laboratory and field part of performance testing was 40 m long.
- Software: The AMS has separate pieces of software for the PLC and the display. The software versions are as follows:
PLC: 3.01R000
Display: TRX_3.01R0000

General remarks

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacturing process for the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. Upon revocation of the publication the certificate loses its validity. After the expiration of the certificate and on request of TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must no longer be used.

The relevant version of this certificate and its expiration date are also accessible on the internet at gal1.de.

Certification of the HM-1400 TRX 2 measuring system is based on the documents listed below and the regular, continuous surveillance of the manufacturer's quality management system:

Initial certification according to EN 15267

Certificate no.:0000059869: 4 September 2018

Expiry date of the certificate: 16 July 2023

Test report: 936/21238805/C dated 10 May 2018

TÜV Rheinland Energy GmbH, Cologne

Publication: BAnz AT 17.07.2018 B9, chapter I number 2.1

UBA announcement dated 3 July 2018

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	DURAG GmbH
AMS designation	HM-1400 TRX 2
Serial number of units under test	755 / 756 / 519 / 520
Measuring principle	Thermocatalytic reaction and cold vapour AAS

Test report

Test laboratory	936/21238805/C TÜV Rheinland
Date of report	2018-05-10

Measured component

Certification range	Hg 0 - 15 µg/m ³
---------------------	--------------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.11 µg/m ³
Sum of negative CS at zero point	-0.19 µg/m ³
Sum of positive CS at span point	0.22 µg/m ³
Sum of negative CS at span point	-0.53 µg/m ³
Maximum sum of cross-sensitivities	-0.53 µg/m ³
Uncertainty of cross-sensitivity	u_i -0.306 µg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u^2
Standard deviation from paired measurements under field conditions *	u_D 0.160 µg/m ³		0.026 (µg/m ³) ²
Lack of fit	u_{lof} -0.081 µg/m ³		0.007 (µg/m ³) ²
Zero drift from field test	$u_{d,z}$ 0.191 µg/m ³		0.036 (µg/m ³) ²
Span drift from field test	$u_{d,s}$ 0.251 µg/m ³		0.063 (µg/m ³) ²
Influence of ambient temperature at span	u_t 0.153 µg/m ³		0.023 (µg/m ³) ²
Influence of supply voltage	u_v 0.078 µg/m ³		0.006 (µg/m ³) ²
Cross-sensitivity (interference)	u_i -0.306 µg/m ³		0.094 (µg/m ³) ²
Influence of sample gas flow	u_p 0.090 µg/m ³		0.008 (µg/m ³) ²
Uncertainty of reference material at 70% of certification range	u_{rm} 0.121 µg/m ³		0.015 (µg/m ³) ²

* The larger value is used :
"Repeatability standard deviation at set point" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)	$u_c = \sqrt{\sum (u_{max, i})^2}$	0.53 µg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	1.03 µg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the ELV 6 µg/m³	17.2
Requirement of EN 15267-3	U in % of the ELV 6 µg/m ³	40.0
	U in % of the ELV 6 µg/m ³	30.0