

CONFIRMATION

Notification: 0000062068_01_01
on changes according to EN 15267 regarding certificate 0000062068_01 dated 05 November 2019

AMS designation: MCS200HW for CO, NO, NO₂, N₂O, SO₂, HCl, NH₃, CH₄, TOC, O₂, H₂O and CO₂

Manufacturer: SICK AG
Nimburger Straße 11
79276 Reute
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 29 June 2021
Publication: BAnz AT 05.08.2021 B5**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**46 Notification as regards Federal Environment Agency (UBA) notice
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter I number 1.4)**


The MCS200HW modular measuring system for CO, NO, NO₂, N₂O, SO₂, HCl, NH₃, CH₄, H₂O, CO₂, TOC and O₂ manufactured by SICK AG can be optionally equipped with an A/C unit.


With an integrated A/C unit, the measuring system can be used in an ambient temperature range from 5°C to 50°C. Without an integrated A/C unit, the measuring system can be used in an ambient temperature range of 5°C to 40°C.

In addition to the basic display, the measuring system can also be equipped with the larger web display on the front panel. The latest software version of the measuring system is 9264565_1.4.2.1R_13FU.

Statement issued by TÜV Rheinland Energy GmbH dated 08 February 2021

TÜV Rheinland Energy GmbH
Cologne, 16 August 2021


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.