

CONFIRMATION

Notification: 0000069259_00_01
on changes according to EN 15267 regarding certificate 0000069259_00 dated 17 June 2020

AMS designation: FP330 for velocity

Manufacturer: Siemens
Östliche Rheinbrückenstr. 50
76187 Karlsruhe
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 09 March 2022
Publication: BAnz AT 11.04.2022 B10**

**VI. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**42 Notification as regards Federal Environment Agency (UBA) notice
of 31 March 2020 (BAnz AT 07.05.2020 B8, chapter I number 2.1)**

The current software version of the measuring device FP330 for exhaust gas velocity
of the company Siemens AG reads:


QAL-1.0.4


The software versions QAL-1.0.0, QAL-1.0.1, QAL-1.0.2 and QAL-1.0.3 are included in
this.

In addition to the previously used housing for panel mounting, the evaluation unit can
also be installed in a wall-mounting housing in the future.

Statement issued by TÜV Rheinland Energy GmbH dated 20 August 2021

TÜV Rheinland Energy GmbH
Cologne, 12 April 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.