

# CERTIFICATE

## of Product Conformity (QAL1)

Certificate No.: 0000072195

**AMS designation:** HORIBA PG-350 P-AMS for CO, NO<sub>x</sub>, CO<sub>2</sub> and O<sub>2</sub>

**Manufacturer:** HORIBA Europe GmbH  
Hans-Mess-Str. 6  
61440 Oberursel  
Germany

**Test Laboratory:** TÜV Rheinland Energy GmbH

**This is to certify that the P-AMS has been tested  
and found to comply with the standards  
EN 15267-1 (2009), EN 15267-2 (2009), EN 15267-4 (2017),  
EN 14793 (2017) and EN 14181 (2014).**

Certification is awarded in respect of the conditions stated in this certificate  
(this certificate contains 9 pages).



Suitability Tested  
EN 15267  
QAL1 Certified  
Regular Surveillance

www.tuv.com  
ID 0000072195

Publication in the German Federal Gazette  
(BAnz) of 31 July 2020

This certificate will expire on:  
30 July 2025

German Federal Environment Agency  
Dessau, 07 September 2020

TÜV Rheinland Energy GmbH  
Cologne, 06 September 2020



Dr. Marcel Langner  
Head of Section II 4.1



ppa. Dr. Peter Wilbring

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Phone: + 49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to certificate D-PL-11120-02-00.

<b>Test Report:</b>	936/21245114/A dated 13 February 2020
<b>Initial certification:</b>	31 July 2020
<b>Expiry date:</b>	30 July 2025
<b>Publication:</b>	BAnz AT 31.07.2020 B10, chapter I number 4.1

### Approved application

The tested P-AMS is suitable for periodic measurements of stationary-source industrial emissions and as a standard reference method for calibrating and validating stationary AMS in the context of QAL2 and AST in accordance with standard EN 14181 at plants according to Directive 2010/75/EU, chapter III (13<sup>th</sup> BImSchV), chapter IV (17<sup>th</sup> BImSchV), 30<sup>th</sup> BImSchV, 44<sup>th</sup> BImSchV, plants in compliance with TA Luft and plants according to the 27<sup>th</sup> BImSchV. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the P-AMS for this application was assessed on the basis of a laboratory test and five field tests at different industrial plants. The plants were two municipal waste incinerators, a lignite-fired power plant, a sewage incinerator and a biomass heating plant.

The AMS is approved for an ambient temperature range of +5 °C to +40 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the limit values and oxygen concentrations relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended purpose.

### Basis of the certification

This certification is based on:

- Test report no. 936/21245114/A dated 13 February 2020 issued by TÜV Rheinland Energy GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process



Publication in the German Federal Gazette: BAnz AT 31.07.2020 B10, chapter I number 4.1,  
UBA announcement of 27 May 2020:

**AMS designation:**

HORIBA PG-350 P-AMS for CO, NO<sub>x</sub>, CO<sub>2</sub> and O<sub>2</sub>

**Manufacturer:**

HORIBA Europe GmbH, Oberursel

**Field of application:**

Portable AMS for periodic measurements of measurements of emissions from stationary sources and as a standard reference method for calibrating and validating stationary AMS in the context of QAL2 and AST in accordance with standard EN 14181 at plants requiring official approval and plants within the scope of the 27<sup>th</sup> and 44<sup>th</sup> BIm-SchV.

**Measuring ranges during performance testing:**

Component	Certification range	supplementary range	Unit
CO	0–75	0–6250	mg/m <sup>3</sup>
NO <sub>x</sub>	0–102,5 <sup>1)</sup>	0–2050 <sup>2)</sup>	mg/m <sup>3</sup>
CO <sub>2</sub>	0–20	-	Vol.-%
O <sub>2</sub>	0–25	0–10	Vol.-%

<sup>1)</sup> as NO<sub>2</sub>, this corresponds to apx. 0–67 mg/m<sup>3</sup> NO

<sup>2)</sup> as NO<sub>2</sub>, this corresponds to apx. 0–1340 mg/m<sup>3</sup> NO

**Software version:**

P20007880001F / 1.19

**Restrictions:**

None

**Note:**

In the event of temperature changes of more than 6 °C, it must be checked on-site whether the measurement uncertainty is still within the permissible limits.

**Test Report:**

TÜV Rheinland Energy GmbH, Cologne

Report no.: 936/21245114/A dated 13 February 2020

### Certified product

This certification applies to automated measurement systems conforming to the following description:

The HORIBA PG-350 P-AMS under test comprises the following components: the PG-350EU analyser, a test gas cooler, a heated sample gas line and a heated sampling probe equipped with a filter.

The analyser measures multiple components and uses component-specific measuring principles. NO<sub>x</sub> is measured using chemiluminescence (CLD), CO and CO<sub>2</sub> are determined with the help of infrared absorption (NDIR) and O<sub>2</sub> is measured with the help of paramagnetism.

The gas to be measured is led to the analyser via a heated probe. A heated sample gas line transports the sample gas to the cooling unit and then via an unheated PFTE line to the analyser.

The PSS-5H cooler is used for gas conditioning and is installed in a plastic case, which also contains the sample gas pump. The sample gas line can be heated electrically to a maximum of 200 °C. The PSP4000-H sampling probe is portable and is equipped with an external ceramic filter element. The probe can be set to a maximum temperature of 180 °C.

The HORIBA PG-350 P-AMS under test consists of the following components:

- PG-350EU multigas analyser,
- Gas conditioning cooler type PSS-5H,
- Sample gas pump,
- Heated sample gas line, max 200 °C, made of PFTE, max. length used during performance testing: 5m, and
- PSP4000-H sampling probe, portable and heated to max. 180 °C, ceramic filter, length during performance testing: 1m.

### General remarks

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacturing process for the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. Upon revocation of the publication the certificate loses its validity. After the expiration of the certificate and on request of TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must no longer be used.

The relevant version of this certificate and its expiration date are also accessible on the internet at [gal1.de](http://gal1.de).



### Document history

Certification of the HORIBA PG-350 P-AMS measuring system is based on the documents listed below and the regular, continuous surveillance of the manufacturer's quality management system:

### Initial certification according to EN 15267

Certificate no.0000072195: 07 September 2020  
Expiry date of the certificate: 30 July 2025  
Test report 936/21245114/A dated 13 February 2020  
TÜV Rheinland Energy GmbH, Cologne  
Publication: BAnz AT 31.07.2020 B10, chapter I number 4.1  
UBA announcement of 27 May 2020

The following pages present the uncertainty calculations for the individual components. Since separate uncertainty calculations are required for each field test, this certificate indicates the highest uncertainty result determined for each field test. All other uncertainty calculations are summarised in the mentioned performance test.

**Calculation of overall uncertainty according to EN 14181 and EN 15267-4 for both systems during field test 3**

**Measuring system**

Manufacturer	HORIBA Europe GmbH
AMS designation	HORIBA PG-350 P-AMS
Serial number of units under test	7DB92A3P / VWG18APN
Measuring principle	Infrared absorption

**Test report**

Test laboratory	936/21245114/A TÜV Rheinland
Date of report	13.02.2020

**Measured component**

Certification range	CO 0 - 75 mg/m <sup>3</sup>
---------------------	--------------------------------

**Evaluation of the cross-sensitivity (CS)**

(system with largest CS)

The cross-sensitivities were calculated site-specifically as a function of the exhaust gas matrix at the respective field test facility, taking into account the cross-sensitivity influences determined in the laboratory.

Maximum sum of cross-sensitivities	0.29 mg/m <sup>3</sup>
Uncertainty of cross-sensitivity	$u_i$ 0.170 mg/m <sup>3</sup>

**Calculation of the combined standard uncertainty**

**Test parameter**

		$u^2$
Standard deviation laboratory test	$u_r$ 0.160 mg/m <sup>3</sup>	0.026 (mg/m <sup>3</sup> ) <sup>2</sup>
Lack of fit	$u_{lof}$ 0.082 mg/m <sup>3</sup>	0.007 (mg/m <sup>3</sup> ) <sup>2</sup>
Zero drift from field test	$u_{d,z}$ 0.260 mg/m <sup>3</sup>	0.068 (mg/m <sup>3</sup> ) <sup>2</sup>
Span drift from field test	$u_{d,s}$ -0.823 mg/m <sup>3</sup>	0.677 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of ambient temperature from field	$u_t$ 0.600 mg/m <sup>3</sup>	0.360 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of supply voltage field test specific	$u_v$ 0,000 mg/m <sup>3</sup>	0.000 (mg/m <sup>3</sup> ) <sup>2</sup>
Cross-sensitivity field test specific	$u_i$ 0.170 mg/m <sup>3</sup>	0.029 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of sample gas flow field test specific	$u_p$ 0,000 mg/m <sup>3</sup>	0.000 (mg/m <sup>3</sup> ) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	$u_{rm}$ 0.606 mg/m <sup>3</sup>	0.368 (mg/m <sup>3</sup> ) <sup>2</sup>

Combined standard uncertainty ( $u_c$ )	$u_c = \sqrt{\sum (u_{max,i})^2}$	1.24 mg/m <sup>3</sup>
Total expanded uncertainty	$U = u_c * k = u_c * 1,96$	2.43 mg/m <sup>3</sup>

**Relative total expanded uncertainty**

<b>Requirement of 2010/75/EU</b>	<b>U in % of the range 50 mg/m<sup>3</sup></b>	<b>4.9</b>
Requirement of EN 15267-3	U in % of the range 50 mg/m <sup>3</sup>	7.5
Requirement for standard reference methods	U in % of the range 50 mg/m <sup>3</sup>	6.0

## Calculation of overall uncertainty according to EN 14181 and EN 15267-4 for both systems during field test 1

### Measuring system

Manufacturer	HORIBA Europe GmbH
AMS designation	HORIBA PG-350 P-AMS
Serial number of units under test	7DB92A3P / VWG18APN
Measuring principle	Infrared absorption

### Test report

Test report	936/21245114/A
Test laboratory	TÜV Rheinland
Date of report	13.02.2020

### Measured component

Measured component	CO <sub>2</sub>
Certification range	0 - 20 Vol.-%

### Evaluation of the cross-sensitivity (CS)

(system with largest CS)

The cross-sensitivities were calculated site-specifically as a function of the exhaust gas matrix at the respective field test facility, taking into account the cross-sensitivity influences determined in the laboratory.

Maximum sum of cross-sensitivities	0.05 Vol.-%
Uncertainty of cross-sensitivity	$u_i$ 0.031 Vol.-%

### Calculation of the combined standard uncertainty

#### Test parameter

				$u^2$
Standard deviation laboratory test	$u_r$	0.020 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Lack of fit	$u_{lof}$	0.081 Vol.-%		0.007 (Vol.-%) <sup>2</sup>
Zero drift from field test	$u_{d,z}$	0.035 Vol.-%		0.001 (Vol.-%) <sup>2</sup>
Span drift from field test	$u_{d,s}$	0.173 Vol.-%		0.030 (Vol.-%) <sup>2</sup>
Influence of ambient temperature from field	$u_t$	0.346 Vol.-%		0.120 (Vol.-%) <sup>2</sup>
Influence of supply voltage field test specific	$u_v$	0.000 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Cross-sensitivity field test specific	$u_i$	0.031 Vol.-%		0.001 (Vol.-%) <sup>2</sup>
Influence of sample gas flow field test specific	$u_b$	0.000 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	$u_{rm}$	0.162 Vol.-%		0.026 (Vol.-%) <sup>2</sup>

Combined standard uncertainty ( $u_c$ )	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.43 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1,96$	0.84 Vol.-%

### Relative total expanded uncertainty

Requirement of 2010/75/EU	<b>U in % of the range 20 Vol.-%</b>	<b>4.2</b>
Requirement of EN 15267-3	<b>U in % of the range 20 Vol.-%</b>	<b>10.0 **</b>
Requirement for standard reference methods	U in % of the range 20 Vol.-%	7.5
	U in % of the range 20 Vol.-%	6.0

\*\* The EU-directive 2010/75/EC on industrial emissions does not requirements for this component.  
A value of 10.0 % was used instead.



**Calculation of overall uncertainty according to EN 14181 and EN 15267-4 for both systems during field test 3**

**Measuring system**

Manufacturer	HORIBA Europe GmbH
AMS designation	HORIBA PG-350 P-AMS
Serial number of units under test	7DB92A3P / VWG18APN
Measuring principle	Chemiluminescence

**Test report**

Test laboratory	936/21245114/A TÜV Rheinland
Date of report	13.02.2020

**Measured component**

Certification range	NO 0 - 67 mg/m <sup>3</sup>
---------------------	--------------------------------

**Evaluation of the cross-sensitivity (CS)**

(system with largest CS)

The cross-sensitivities were calculated site-specifically as a function of the exhaust gas matrix at the respective field test facility, taking into account the cross-sensitivity influences determined in the laboratory.

Maximum sum of cross-sensitivities	0.31 mg/m <sup>3</sup>
Uncertainty of cross-sensitivity	$u_i$ 0.179 mg/m <sup>3</sup>

**Calculation of the combined standard uncertainty**

**Test parameter**

		$u^2$
Standard deviation laboratory test	$u_r$ 0.100 mg/m <sup>3</sup>	0.010 (mg/m <sup>3</sup> ) <sup>2</sup>
Lack of fit	$u_{lof}$ -0.348 mg/m <sup>3</sup>	0.121 (mg/m <sup>3</sup> ) <sup>2</sup>
Zero drift from field test	$u_{d,z}$ 0.039 mg/m <sup>3</sup>	0.002 (mg/m <sup>3</sup> ) <sup>2</sup>
Span drift from field test	$u_{d,s}$ -1.006 mg/m <sup>3</sup>	1.012 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of ambient temperature from field	$u_t$ 0.663 mg/m <sup>3</sup>	0.440 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of supply voltage field test specific	$u_v$ 0,000 mg/m <sup>3</sup>	0.000 (mg/m <sup>3</sup> ) <sup>2</sup>
Cross-sensitivity field test specific	$u_i$ 0.179 mg/m <sup>3</sup>	0.032 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of sample gas flow field test specific	$u_p$ 0,000 mg/m <sup>3</sup>	0.000 (mg/m <sup>3</sup> ) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	$u_{rm}$ 0.542 mg/m <sup>3</sup>	0.293 (mg/m <sup>3</sup> ) <sup>2</sup>
Converter efficiency for AMS measuring NOx	$u_{ce}$ 1.277 mg/m <sup>3</sup>	1.630 (mg/m <sup>3</sup> ) <sup>2</sup>

Combined standard uncertainty ( $u_c$ )

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 1.88 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1,96 \quad 3.69 \text{ mg/m}^3$$

**Relative total expanded uncertainty**

**U in % of the range 45 mg/m<sup>3</sup> 8.2**

**Requirement of 2010/75/EU**

**U in % of the range 45 mg/m<sup>3</sup> 20.0**

Requirement of EN 15267-3

U in % of the range 45 mg/m<sup>3</sup> 15.0

Requirement for standard reference methods

U in % of the range 45 mg/m<sup>3</sup> 10.0



## Calculation of overall uncertainty according to EN 14181 and EN 15267-4 for both systems during field test 1

### Measuring system

Manufacturer	HORIBA Europe GmbH
AMS designation	HORIBA PG-350 P-AMS
Serial number of units under test	7DB92A3P / VWG18APN
Measuring principle	Paramagnetic

### Test report

Test laboratory	936/21245114/A
Date of report	TÜV Rheinland
	13.02.2020

### Measured component

Certification range	O <sub>2</sub>
	0 - 25 Vol.-%

### Evaluation of the cross-sensitivity (CS)

(system with largest CS)

The cross-sensitivities were calculated site-specifically as a function of the exhaust gas matrix at the respective field test facility, taking into account the cross-sensitivity influences determined in the laboratory.

Maximum sum of cross-sensitivities	0.00 Vol.-%
Uncertainty of cross-sensitivity	u <sub>i</sub> 0.000 Vol.-%

### Calculation of the combined standard uncertainty

#### Test parameter

				u <sup>2</sup>
Standard deviation laboratory test	u <sub>r</sub>	0.010 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Lack of fit	u <sub>lof</sub>	0.016 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Zero drift from field test	u <sub>d,z</sub>	-0.017 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Span drift from field test	u <sub>d,s</sub>	-0.046 Vol.-%		0.002 (Vol.-%) <sup>2</sup>
Influence of ambient temperature from field	u <sub>t</sub>	0.346 Vol.-%		0.120 (Vol.-%) <sup>2</sup>
Influence of supply voltage field test specific	u <sub>v</sub>	0.000 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Cross-sensitivity field test specific	u <sub>i</sub>	0.000 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Influence of sample gas flow field test specific	u <sub>b</sub>	0.000 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	u <sub>rm</sub>	0.202 Vol.-%		0.041 (Vol.-%) <sup>2</sup>

Combined standard uncertainty (u <sub>c</sub> )	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.40 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1,96$	0.79 Vol.-%

### Relative total expanded uncertainty

Requirement of 2010/75/EU	<b>U in % of the range 25 Vol.-%</b>	<b>3.2</b>
Requirement of EN 15267-3	<b>U in % of the range 25 Vol.-%</b>	<b>10.0 **</b>
Requirement for standard reference methods	U in % of the range 25 Vol.-%	7.5
	U in % of the range 25 Vol.-%	6.0

\*\* The EU-directive 2010/75/EC on industrial emissions does not requirements for this component.  
A value of 10.0 % was used instead.