

CONFIRMATION

Notification: 0000072196_01_03
on changes according to EN 15267 regarding certificate 0000072196_01 dated 02 June 2021

AMS designation: MGAprime Q for CO, NO, NO₂, N₂O, SO₂, O₂ and CO₂

Manufacturer: MRU GmbH
Fuchshalde 8
74172 Neckarsulm
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 05 July 2023
Publication: BAnz AT 02.08.2023 B7**

III. Notifications as regards the uniform practice in monitoring emissions and ambient air:

13 Notification as regards Federal Environment Agency (UBA) notices of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter I number 4.1) and of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 17)


The current software version of the MGAprimeQ portable measuring system for CO, NO,
NO₂, SO₂, N₂O, CO₂ and O₂ from MRU GmbH is:
V1.003.008.


Furthermore, the intermediate versions V1.002.009; V1.003.000; V1.003.001;
V1.003.002; V1.003.004; V1.003.005; V1.003.006; V1.003.007 can be used.

The ambient temperature sensor can also be equipped with an angled connector. The
upper part of the gas cooler has been modified by means of an additional pressure plate
in the area of the screw connections.

Statement issued by TÜV Rheinland Energy GmbH dated 08 February 2023

TÜV Rheinland Energy GmbH
Cologne, 02 August 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.